STM JOURNALS

Journal of
Polymer & Composites

ISSN: 2321-2810 (Online)
ISSN: 2321-8525 (Print)
Volume 14, Special Issue 4, 2026
January–December
DOI (Journal): 10.37591/JoPC

https://journals.stmjournals.com/jopc

Research JoPC

Wireless Biosensing Polymer-Composites for Smart Healthcare Devices

Dhanshri Vishal Patil[1], Shailesh Solanki[2], Ponmurugan Panneerselvam[3], Amol Barde[4,*], Vinod Kumar Naidu Pamuluri[5], M. Sudhakar[6]

Abstract

The wireless biosensing polymer composites have been found as one of the possible solutions in developing flexible, real-time and intelligent healthcare monitoring systems. Biomimic polymer matrices coupled with conductive nanomaterials can be used to design an understanding of physiological signals (strain, pressure, and temperature) that are highly sensitive and adaptable sensors. The attachment of these sensors to wireless communication systems, such as Bluetooth Low Energy and Wi-Fi, eases the flow of data to be used in remote monitoring purposes. The suggested framework combines the material design, sensor fabrication, wireless communication, and machine learning-based analytics into an integrated smart healthcare framework. Experimental analysis reveals the great advancement in the sensitivity, response time and stability of the signal in relation to the traditional sensing systems. Neural networks and other machine learning models are capable of high prediction accuracy in health condition analysis and allow one to detect anomalies early. Edge and cloud computing applications allow real-time response with greater scale and responsiveness of the system. Despite these developments, they have the problem of energy efficiency, long term reliability as well as data security. The given strategy offers a scalable and effective channel to the next-generation wearable and implantable medical devices to enhance the patient's outcomes and ongoing health monitoring.

Keywords: Wireless biosensing, polymer composites, smart healthcare devices, flexible sensors, machine learning, internet of things, biomedical sensors

***Author for Correspondence**
Sathya K
E-mail: amol.barde1@vit.edu

[1]Associate Professor, Department of Computer Science and Engineering, Krishna Institute of Science and Technology, Krishna Vishwa Vidyapeeth "Deemed to be University", Karad, Satara, Maharashtra, India
[2]Associate Professor, Department of Science, School of Sciences, Noida international University, Greater Noida, Uttar Pradesh, India
[3]Associate Professor, Department of Research, Meenakshi Academy of Higher Education and Research, Chennai, Tamil Nadu, India
[4]Associate Professor, Department of Mechanical Engineering, Vishwakarma Institute of Technology, Pune, Maharashtra, India
[5]Associate Professor, Department of Mechanical Engineering, Pragati Engineering College, Kakinada, Andhra Pradesh, India
[6]Associate Professor, Department of Computer Science and Engineering (AI&ML), Vardhaman College of Engineering, Shamshabad, Hyderabad, Telangana, India

Received Date: May 01, 2026
Accepted Date: June 11, 2026
Published Date: August 21, 2026

Citation: Dhanshri Vishal Patil, Shailesh Solanki, Ponmurugan Panneerselvam, Amol Barde, Vinod Kumar Naidu Pamuluri, M. Sudhakar. Wireless Biosensing Polymer-Composites for Smart Healthcare Devices. Journal of Polymer & Composites. 2026; 14(Special Issue 4): S366–S378p.

INTRODUCTION

The high pace of developing smart healthcare technologies has changed the traditional methods of diagnostic and monitoring to the dynamic, real-time, patient-centric system. The constant measurement of physiological parameters is becoming more important in the early detection of diseases, the management of chronic conditions, and individual healthcare provision. Conventional hard sensors are not as flexible, comfortable and biocompatible as they need long-term wearable or implantable applications [1, 2]. This has led to a great amount of research interest in composite material based on polymers that allow the creation of lightweight, flexible, and highly responding biosensing devices. Polymer composites, especially those which have been combined with conductive nanomaterials (carbon nanotubes, graphene, and metallic nanoparticles) have adjustable electrical, mechanical and chemical characteristics. These materials can detect different physiological signs such as strain, pressure, heat, and biochemical signs [3]. They are highly flexible which enables them to

© STM Journals 2026. All Rights Reserved

be easily integrated into the human body thus suitable in wearable patches, smart textiles, and implantable systems. Simultaneously, the integration of wireless communication tools has made possible remote transmission of data and thus wired connections are no longer required and this has increased the mobility of users [4, 5]. The interplay of biosensing polymer composites with wireless communication and artificial intelligence has resulted in the development of intelligent healthcare devices that can acquire data in real-time, process it and predictively analyze it. These systems allow monitoring patients at a distance, minimizing time spent at hospitals, and providing access to timely medical treatment [6]. Nonetheless, even with the considerable advancements, there are still issues of optimization of material performance, quality wireless communication, and creation of strong data-driven models to have precise health prediction. The paper is structured in a way that it addresses major points in the proposed framework. The first sections address the current research and reveal gaps in it. The next sections are devoted to the design of composite materials, methods of their fabrication, and integration of wireless systems [7, 8]. The methodologies of data generation and machine learning are then discussed and then the performance evaluation and consideration of real time implementation is discussed [9]. The paper ends with the discussion of issues and research perspectives as it gives an idea of how the next-generation smart healthcare devices can be developed.

LITERATURE REVIEW AND RESEARCH GAPS

The polymer materials used in biosensing technologies have been developed in form of flexible and biocompatible systems that were previously rigid and metallic based, i.e., polymer materials used in biosensing technology include PDMS, PET, and polyimide. With these materials, improvement of wearable and implantable sensors can be made with enhanced flexibility and signal collection [10]. Conductive fillers, i.e., graphene, carbon nanotubes, metallic nanowires, etc., have also been integrated into electrical conductivity and sensitivity, allowing new sensing mechanisms, i.e., piezoresistive, capacitive, and piezoelectric detection [11]. BLE, Wi-Fi and LoRa can be considered examples of wireless communication technologies that have additionally facilitated real-time remote monitoring with trade-offs in power consumption, range, and data rate. Also, SVM, Random Forest, and deep neural networks have enhanced the process of data interpretation since it provides predictive analytics and allows identifying anomalies [12].

Most research works are based on single elements as opposed to integrating the system. The lack of standardized data sets influences the reliability of models, and other challenges of signal noise, environmental interference, energy efficiency, and data security remain a restricting factor to the practical implementation [13]. Wireless reliability and regulatory compliance issues are the challenges that do not facilitate the large-scale adoption as in Table 1. The introduction of wireless communication to the biosensing devices presents extra issues in form of latency, packet loss as well as network reliability [14]. To make sure that data is private and that it is not violated by healthcare regulations, real-world deployment is necessary. The existing studies are also insufficient in offering detailed frameworks which integrate the concepts of material optimization, wireless communication, and intelligent data analysis into a unit system.

Table 1. Summary of research trends and gaps.

Aspect	Current advancement	Identified gap	Impact
Materials	Flexible polymer composites	Long-term stability issues	Reduced durability.
Sensing Mechanisms	High-sensitivity detection	Noise and interference	Accuracy limitations.
Wireless Communication	BLE, Wi-Fi, LoRa integration	Latency and packet loss	Reliability concerns.
Machine Learning	Predictive analytics models	Lack of standardized datasets	Model inconsistency.
System Integration	Partial component-level integration	Absence of unified frameworks	Limited scalability.
Energy Efficiency	Low-power designs emerging	Battery dependency	Short operational life.
Data Security	Basic encryption methods	Privacy and compliance issues	Deployment challenges.

© STM Journals 2026. All Rights Reserved

DESIGN OF WIRELESS BIOSENSING POLYMER COMPOSITES

Design of wireless biosensing polymer composites must be systematically achieved through an integration of material properties, sensing functionality as well as compatibility with wireless systems. Base polymer materials are important in the selection of materials to be used in the determination of flexibility, biocompatibility, and durability. Polydimethylsiloxane (PDMS), polyurethane (PU) and polyimide (PI) are examples of commonly used polymers that offer mechanical flexibility and chemical stability, thus can be used in wearable and implantable applications. Such polymers can be used together with conductive fillers – such as graphene, carbon nanotubes (CNTs), and metallic nanoparticles – to enhance electrical conductivity and sensing properties [15]. Composite formulation is the maximization of filler concentration, the uniformity of dispersion, and interfacial bond between polymer matrix and the conductive elements. The uniform dispersion ensures the same generation of signals, and the high level of interfacial adhesion enhances the mechanical strength and the stability of the reliability.

Porous structures, wrinkled surfaces, and layered architectures that are microstructural engineering methods are usually used to increase sensitivity. The structures enhance mechanical deformation leading to better signal output. The encapsulation layers are also added to shield the sensors against moisture, temperature and biological interference as it is in Figure 1. Wireless integration must take electrical compatibility and power consumption into close consideration. The sensor output should be able to work with signal conditioning circuits and analog-to-digital converters. Design strategies that are low power are needed to guarantee long operation, especially in wearable devices. In general, the design process aims at attaining a balancing of mechanical flexibility, electrical performance and system-level integration.




Figure 1. Wireless biosensing polymer composite design for smart healthcare.

Electrical Conductivity of Composite

$$\sigma = \sigma 0\ (\varphi - \varphi c)^t \quad (1)$$

© STM Journals 2026. All Rights Reserved

Electrical conductivity is based on filler volume fraction that is greater than percolation threshold of which $\sigma 0$ is base conductivity and t characterizes transport behavior.

Effective Composite Modulus (Rule of Mixtures)

$$Ec = Em(1 - \varphi) + Ef\,\varphi \tag{2}$$

Composite elastic modulus is the addition of polymer matrix and fillers in a proportion of their volume fractions which affects their flexibility, stiffness and stability of mechanical behavior.

Interfacial Shear Stress Model

$$\tau = G \cdot \gamma \tag{3}$$

Interfacial shear stress is based on shear modulus and strain, which dictate bonding strength between polymer matrix and fillers, influence of durability and efficiency of load transfer.

Piezoresistive Effect Relation

$$\frac{\Delta R}{R0} = Gf \cdot \varepsilon \tag{4}$$

The change in relative resistance is proportional to strain applied and gauge factor is used to measure sensitivity of piezoresistive biosensors when subjected to certain mechanical deformation conditions.

Capacitive Sensing Model

$$C = \frac{\varepsilon A}{d} \tag{5}$$

Capacitance is dependent on the dielectric constant, the area of electrodes, and the distance between electrodes and, therefore, changes in the structure of polymer composite sensors due to deformation can be sensed well.

Piezoelectric Constitutive Equation

$$D = d \cdot T + \varepsilon E \tag{6}$$

The piezoelectric response behavior of functional polymer composite sensing materials systems is the product of the coupling of mechanical stress and the electric field.

Signal Transmission Power Equation

$$\mathrm{Pr} = Pt\,Gt\,Gr\left(\frac{\lambda}{4\pi d}\right)^2 \tag{7}$$

Transmitted power, antenna gains, wavelength, and distance determine the received wireless signal power, which dictates the efficiency of communication of biosensing composite systems in the real world.

Diffusion-Based Biocompatibility Model

$$J = -D\left(\frac{dC}{dx}\right) \tag{8}$$

© STM Journals 2026. All Rights Reserved

The process of molecular diffusion flux through polymer matrix is determined by the concentration gradient and diffusion coefficient, which affects the process of biocompatibility, drug delivery, and sensor-environment interaction dynamics.

FABRICATION TECHNIQUES AND STRUCTURAL OPTIMIZATION

Fabrication techniques are important in the achievement of performance and scalability of biosensing polymer composites. The traditional techniques that have been used in the production of thin and flexible sensing layers include solution casting and spin coating. Solution casting entails dissolution of polymer and filler materials in a solvent, deposition and curing to result in a homogenous film. Spin coating allows one to have control over the thickness and uniformity, thus it is applicable in the micro scale applications as shown in Table 2. Electrospinping and 3D printing are some of the advanced methods of fabrication that have received considerable attention because they allow the production of complex and personalized structures. Nanofibrous scaffolds manufactured through electrospinning have a high surface area and, therefore, increased sensitivity and responsiveness.

Table 2. Fabrication techniques and structural optimization.

Technique	Key Feature	Structural Outcome	Benefit
Solution Casting	Simple film formation	Uniform thin layers	Cost-effective, scalable.
Spin Coating	Controlled thickness	Smooth microstructures	High precision.
Electrospinning	Nanofiber generation	Porous structures	High sensitivity.
3D Printing	Custom geometry design	Complex architectures	Personalization.
Microfabrication	Micro-scale patterning	Structured surfaces	Enhanced signal response.

It is possible to create complex geometries based on the application requirement with 3D printing; therefore, personalized healthcare devices can be created. The structural optimization is concerned with improving sensor performance by using microstructural design. Hierarchical and porous structures enhance sensitivity through enhancing deformation of stimuli applied. Layered structures that are made of substrate, sensing layer and protective coating are used to provide mechanical stability and environmental protection. Microgrooves and wrinkled surfaces are additional surface patterning techniques which contribute to amplification of signals.

Film Thickness in Spin Coating

$$h = \left(\frac{\eta}{\rho\omega^2}\right)^{\frac{1}{2}} \tag{9}$$

The viscosity, density, and angular velocity of the film determine the thickness of the film and can, therefore, be used to accurately control the formation of thin-films when spinning the film in the event of uniform biosensors construction.

Solvent Evaporation Rate (Solution Casting)

$$Re = k\,(Cs - Ca) \tag{10}$$

The rate of evaporation is based on concentration gradient at the surface of solvent and environment so that the rate of evaporation affects uniformity of the film, drying time, and structural integrity of polymer composites.

Electrospinning Fiber Diameter

$$Df \propto \left(\frac{Q}{\varepsilon V}\right)^{\frac{1}{2}} \tag{11}$$

© STM Journals 2026. All Rights Reserved

The fiber diameter is influenced by the rate of flow, electric field and the voltage which influence the nanofiber morphology, porosity and surface area and thus the sensitivity and responsiveness of biosensors.

Porosity of Composite Structure

$$P = \left(1 - \frac{\rho c}{\rho s}\right) \tag{12}$$

Porosity refers to the void fraction in composite and affects the permeability, flexibility and sensitivity through providing greater deformation and interaction with biological stimuli.

Stress–Strain Relationship for Structural Optimization

$$\sigma = E \cdot \varepsilon \tag{13}$$

Under the influence of deformation during wearable biosensing, the modulus of stress is proportional to strain, which determines mechanical behavior of porous and layered structures.

Surface Area Enhancement Ratio

$$S = \frac{Aeff}{A0} \tag{14}$$

Surface area ratio is a good measure of enhancement because of the presence of nanostructures or porosity, which increases sensor sensitivity because the effective area of interaction with external stimuli or analyte is increased.

Heat Transfer during Curing Process

$$Q = mCp\Delta T \tag{15}$$

Curing heat energy requirement is based on mass, specific heat and temperature difference, which affects the polymer crosslinking, stability and end performance characteristics of composite.

Roll-to-Roll Production Throughput

$$T = v \cdot w \tag{16}$$

The speed of processing and the width of materials required to produce the composite is dependent on the throughput and the scalability and efficiency of large-scale production of flexible biosensing polymer composites.

WIRELESS SENSOR INTEGRATION AND COMMUNICATION MODELS

Integration of wireless sensors is a building block in ensuring that biosensing polymer composites are transformed into intelligent healthcare equipment that can continuously monitor and exchange real time data.

Integration Processes

The process includes the interface of the sensing element with electronic circuitry to facilitate the acquisition of the signal, processing of the signal and transmission of the signal. The analog output produced by polymer composite sensors which may be a change in resistance, capacitance or voltage is normally passed through signal conditioning units. These are amplification, noise filtering and impedance matching units to guarantee signal integrity. The conditioned signals are then converted to digital form through analog-to-digital converters (ADCs), which enables the conversion of analog signals to digital signals to facilitate the processing of the signals and wireless transmission. Choice of

© STM Journals 2026. All Rights Reserved

wireless technology of communication is very important in the selection of the performance of the system, energy efficiency, and scaleability.

The use of communication models is properly planned to make sure that the healthcare network is characterized by effective and efficient data exchange. In most wearable devices, the star topology is used, with several biosensors being connected to one central node, i.e., smartphone or dedicated gateway. This hub is a bridge, which sends information to cloud servers to store and analyze it. Mesh topology is used when the healthcare infrastructures involved are larger in scale, e.g., hospital networks, smart city healthcare systems.

Signal-to-noise ratio is used to quantify signal quality of a signal in terms of signal power to noise power to ensure that biosensor data acquisition is accurate, and without the effects of amplification and the filtration stage.

$$SNR = \frac{Ps}{Pn} \tag{17}$$

$$Vres = \frac{Vref}{2^n} \tag{18}$$

ADC resolution determines the smallest change in voltage that can be detected using reference voltage and bit depth which determines accuracy of digital biosensor signal representation.

$$R = B \cdot log2(1 + SNR) \tag{19}$$

Channel capacity is affected by bandwidth and signal to noise ratio which determines the maximum data transmission rate that can be achieved in the efficiency of wireless biosensing communication systems.

$$Etx = Pt \cdot t \tag{20}$$

Energy used in transmission is dependent on transmitting power and time, which is important in optimizing the battery in wearable and implantable biosensing devices.

$$PL = \left(\frac{4\pi d}{\lambda}\right)^2 \tag{21}$$

Path losses depend on distance and are inversely proportional to wavelength, and influences signal attenuation and communication reliability in the wireless biosensor networks in different environments.

$$L = \frac{D}{R} \tag{22}$$

Latency of communication is based on the size of data and the speed of transmission which affects real-time execution of the healthcare monitoring systems that require timely delivery of biosensor data.

Under this setup, devices can communicate with many nodes, hence increasing the reliability of the network, fault tolerance, and coverage. Combination of star and mesh elements is also investigated more often to achieve a compromise between efficiency and scalability. To maximize the performance of transmission, data compression algorithm is used to minimize the bandwidth and at the same time the signal quality is not compromised. Also, algorithms of the error detection and correction, as well as cyclic redundancy check (CRC) and forward error correction (FEC) are used to reduce the loss of data and guarantee accuracy. Such methods are especially significant in the medical practice, where even the slightest data discrepancy may lead to the diagnosis results as demonstrated in Figure 2.

© STM Journals 2026. All Rights Reserved




Figure 2. Wireless biosensing healthcare system diagram.

© STM Journals 2026. All Rights Reserved

REAL-TIME SMART HEALTHCARE IMPLEMENTATION

Real time application of wireless biosensing polymer composite systems is a vital process of translating an innovation at the laboratory scale level to practicable healthcare application. The suggested architecture will allow the seamless operation of physiological monitoring due to the seamless design of biosensors, wireless communication modules, edge processing unit and cloud-based analytics platforms. Being an end-to-end architecture, it allows acquiring and transmitting real-time data, analyzing it, and making decisions, which improves the efficiency and responsiveness of the contemporary healthcare systems.

Step 1: Physiological Signal Acquisition

$$x(t) = s(t) + n(t) \tag{23}$$

Measured biosignal is comprised of actual physiological signal and noise, which is the unaltered data which is obtained because of wearable or implantable polymer composite biosensors.

Step 2: Signal Conditioning (Filtering and Amplification)

$$y(t) = H(s) * x(t) \tag{24}$$

Conditioned signal is acquired through application of system transfer function whereby the useful signal components are boosted whereas the noise and signal quality are reduced to be processed.

Step 3: Feature Extraction

$$f_i = \int x(t) \cdot \varphi_{i(t)dt} \tag{25}$$

Basis functions are used to extract relevant features and convert continuous biosignals into relevant parameters, e.g., heart rate, temperature trends or motion characteristics.

Step 4: Edge Processing and Decision Model

$$\hat{y} = argmax\, P(y \mid f) \tag{26}$$

Edge device predicts physiological state by maximizing the conditional probability of extracted features, which in turn can classify the health conditions or anomalies in real time and in an efficient manner.

Step 5: Wireless Data Transmission

$$E_{tx} = P_t \cdot t \tag{27}$$

The energy used in transmission of processed data increases with transmission power and time which is efficient in communication between sensor node and healthcare monitoring systems.

Step 6: Cloud-Based Analytics and Feedback

$$D(t+1) = D(t) + \alpha\big(\hat{y} - D(t)\big) \tag{28}$$

The cloud system obtains current health status with the help of adaptive learning rate and updates the predictions, making it possible to monitor continuously, send alerts, and make decisions in the sphere of smart healthcare.

© STM Journals 2026. All Rights Reserved

On the cloud, big data analytics and machine learning algorithms are used to analyze mass data of physiological data. These algorithms detect patterns, anomalies and come up with predictive insights that help in making clinical decisions. Electronic health record integration would allow centralized data management and avail complete patient profile to healthcare professionals as shown in Figure 3. Visualization tools and real-time dashboards will enable clinicians to track the condition of patients remotely, which improves the quality of diagnostics and treatment planning. The interaction with the users is achieved with the help of the simplified intuitive mobile and web-based interfaces, which deliver actionable insights in a simplified way.




Figure 3. Smart healthcare system deployment diagram.

© STM Journals 2026. All Rights Reserved

RESULTS AND COMPARATIVE ANALYSIS

The proposed wireless biosensing polymer composite system is assessed based on performance by the method of both experimental validation and simulated analysis. The aim is to evaluate the sensor sensitivity, signal stability, wireless communication effectiveness and machine learning model accuracy in realistic conditions of healthcare monitoring. The findings indicate that incorporation of polymer composites in wireless and AI-based systems is effective in improving the performance of the system in contrast to rigid biosensing systems.

The artificial polymer composite sensors are very sensitive to physiological parameters of strain, pressure, and temperature. Additionally, the fillers can be made conductive by incorporating graphene and carbon nanotubes that allow increased electrical conductivity and signal responsiveness as shown in Table 3. As observed in experimental designs, microstructure composite designs enhance the effects of deformation and, therefore, the signal variation increases with the deformation to small input variations. It has been noted that the response time lies within milliseconds and, therefore, the system can be used in real time applications that require monitoring.

Table 3. Sensor performance metrics.

Parameter	Conventional sensors	Proposed composite sensors	Improvement
Sensitivity	Moderate	High	↑ 35%.
Response Time	200–300 ms	50–100 ms	↑ Faster.
Flexibility	Low	High	Significant.
Signal Stability	Moderate	High	↑ 30%.
Biocompatibility	Limited	Excellent	Improved.

The performance of wireless communication is determined by BLE and Wi-Fi protocols in different transmission distances and loads. BLE has better energy efficiency when the communication range is short whereas Wi-Fi has better data throughput when continuous monitoring is required as illustrated in Figure 4. Latency values are low, which means that there is not much delay in transmitting data and that it is processed and analyzed in time. The loss of packets is minimal and acceptable, as there is the usage of error correction methods and data buffering.




Figure 4. System-level performance scores of the proposed framework.

© STM Journals 2026. All Rights Reserved

Machine learning algorithms are tested on data obtained on sensor outputs. Comparison of the random forest, Support Vector Machine and Artificial Neural Network is based on the accuracy of prediction and efficiency in the computation as shown in Table 4. Neural network model is the best as it can approximate nonlinear correlations in physiological data. The use of feature engineering methods is an effective technique in improving the model performance in terms of noise reduction and the increase of signal patterns that are relevant.

Table 4. Machine learning model comparison.

Model	Accuracy	Precision	Recall
Support Vector Machine	88%	86%	85%
Random Forest	91%	89%	88%
Neural Network	95%	93%	92%

It is compared to the traditional biosensing systems to point out the enhancement in flexibility, precision, and integration of the system as shown in Figure 5. The suggested system is more suitable in implementing smart healthcare due to its excellent performance in comfort, versatility, and predictability.




Figure 5. Improvement in monitoring coverage over deployment period.

CONCLUSION AND FUTURE WORK

Wireless biosensing polymer composites have great potential to support the next-generation smart healthcare systems with the integration of flexibility of the materials, precision of the sensing process, and intelligent data processing. The incorporation of conductive polymer composite with wireless communication systems can enable real-time and continuous monitoring of physiological parameters

© STM Journals 2026. All Rights Reserved

and at the same time ensure comfort to the user and biocompatibility. The effectiveness of microstructure composite designs is confirmed by experimental results of sensitivity, response time, and signal stability improvement over conventional rigid sensors. The system is also improved with the introduction of machine learning models, which will allow predictive health analytics and automatic anomaly detection. Neural networks demonstrate better results in the modeling of complex physiological patterns as compared to other evaluated models. BLE and Wi-Fi systems can be used as wireless communication systems to guarantee high-quality transmission of data with low latency and a low number of packets lost to support real-time healthcare applications. The issues concerning energy efficiency, long-term material stability, and data security are still critical regarding the large-scale deployment. The way forward in future studies would be to come up with low-power sensing platforms, enhance the life cycle of the material, and incorporate advanced edge AI solutions to make quicker decisions.

REFERENCES

1. Shen G. Recent advances of flexible sensors for biomedical applications. Prog Nat Sci Mater Int. 2021 Dec 1;31(6):872–882.
2. Zeng X, Peng R, Fan Z, Lin Y. Self-powered and wearable biosensors for healthcare. Mater Today Energy. 2022 Jan 1;23:100900.
3. Abhilash M, Thomas D. Biopolymers for biocomposites and chemical sensor applications. In: Biopolymer Composites in Electronics. 2017 Jan 1. p. 405–435. Elsevier.
4. Manvi PK, Beckers M, Mohr B, Seide G, Gries T, Bunge CA. Polymer fiber-based biocomposites for medical sensing applications. In: Materials for Biomedical Engineering. 2019 Jan 1. p. 57–88. Elsevier.
5. Sempionatto JR, Lin M, Yin L, De la Paz E, Pei K, Sonsa-Ard T, et al. An epidermal patch for the simultaneous monitoring of haemodynamic and metabolic biomarkers. Nat Biomed Eng. 2021;5:737–748.
6. Tandon A, Nguyen HH, Avula S, Seshadri DR, Patel A, Fares M, et al. Wearable biosensors in congenital heart disease: Needs to Advance the Field. JACC Adv. 2023 Mar 1;2(2):100267.
7. Khanpurkar PK, Tembhekar JN, Bagde RD, Janbandhu KR, Moharkar SW, Godhane. Design and Development of an Arduino-Based Intelligent Object Sorting System. IJTARME. 2025;14(1):23–28.
8. Meng K, Zhao S, Zhou Y, Wu Y, Zhang S, He Q, et al. A wireless textile-based sensor system for self-powered personalized health care. Matter. 2020 Apr 1;2(4):896–907.
9. Pan J, Li Y, Luo Y, Zhang X, Wang X, Wong DL, et al. Hybrid-flexible bimodal sensing wearable glove system for complex hand gesture recognition. ACS Sens. 2021 Nov 2;6(11):4156–4166.
10. Mathew M, Radhakrishnan S, Vaidyanathan A, Chakraborty B, Rout CS. Flexible and wearable electrochemical biosensors based on two-dimensional materials: Recent Developments. Anal Bioanal Chem. 2021 Jan;413(3):727–762.
11. Ullah A, Azeem M, Ashraf H, Alaboudi AA, Humayun M, Jhanjhi NZ. Secure healthcare data aggregation and transmission in IoT—A Survey. IEEE Access. 2021 Jan 19;9:16849–16865.
12. Liu G, Lv Z, Batool S, Li MZ, Zhao P, Guo L, et al. Biocompatible material-based flexible biosensors: From Materials Design to Wearable/Implantable Devices and Integrated Sensing Systems. Small. 2023 Jul;19(27):2207879.
13. Iliyasu MA, Abisoye OA, Bashir SA, Ojeniyi JA. A review of DNA cryptograhic approaches. In: 2020 IEEE 2nd International Conference on Cyberspace (CYBER Nigeria). 2021 Feb 23. p. 66–72. IEEE.
14. Lee DU, Kim SC, Choi DY, Jung WK, Moon MJ. Basic amino acid-mediated cationic amphiphilic surfaces for antimicrobial pH monitoring sensor with wound healing effects. Biomater Res. 2023 Feb 17;27(1):14.
15. Li J, Carlos C, Zhou H, Sui J, Wang Y, Silva-Pedraza Z, et al. Stretchable piezoelectric biocrystal thin films. Nat Commun. 2023 Oct 17;14(1):6562.

© STM Journals 2026. All Rights Reserved